3T MRI PROTOCOLS

Protocol Name: **BRACHIAL PLEXUS - Sadler** Body Part: **BRACHIAL PLEXUS** Coil: **HEAD / NECK ARRAY**

Sequences	Seq Name	Image Plane	# Slices	TR	TE	Slice Thick	FOV	NSA	Matrix	Phase Dir
3 Plane Loc	gre	3 Plane	18	7.8	3.69	6	400	1	256	A>P
Cor T1 FS Large FOV Pre	tse	Coronal	48	845	9.4	3	360	2	256	R>L
Cor STIR Large FOV	tse	Coronal	48	3000	31	3	360	1	256	R>L
Axial T1 Large FOV	tse	Axial	36	634	9.4	3	360	2	256	R>L
Axial STIR Large FOV	tse	Axial	36	5000	96	3	360	2	256	R>L
Axial T1 sm FOV	tse	Axial	48	500	11	3	240	1	256	A>P
Axial T2 FS sm FOV	tse	Axial	48	4190	53	3	240	3	256	A>P
Cor T1 sm FOV	tse	Coronal	38	710	9.8	3	240	2	256	R>L
Cor STIR	tse	Coronal	38	3520	51	3	240	1	256	R>L
Sag STIR	tse	Sagittal	50	3520	51	3.5	240	2	256	A>P
Gadolinium										
Sag T1 FS Post	tse	Sagittal	50	590	11	3.5	240	1	256	A>P
Axial T1 FS Sm FOV Post	tse	Axial	36	630	9.6	3	240	1	256	R>L
Axial T1 FS Post	tse	Axial	36	634	9.4	3	360	2	256	R>L
Cor T1 FS Post	tse	Coronal	48	650	9.4	3	360	1	256	R>L